

CIPO
CANADIAN INTELLECTUAL
PROPERTY OFFICE

Ottawa Hull K1A 0C9

(21) (A1) 2,169,803 (22) 1996/02/19 (43) 1996/11/20

- 6
 (51) Int.Cl. GO1N 33/96; GO1N 33/49; A61B 6/00; A61B 5/14
- (19) (CA) APPLICATION FOR CANADIAN PATENT (12)
- (54) Method to Prepare Dye-Based Reference Material
- (72) Li, Jay J. U.S.A.;
 Davey, Justin E. U.S.A.;
 Swist, David P. U.S.A.;
 Voo, Liann U.S.A.;
- (71) Ciba Corning Diagnostics Corp. U.S.A.
- (30) (US) 08/447,168 1995/05/19
- (57) 23 Claims

Non 3. Logical (in imbrant to dams)
(him 6. abstract

Notice: This application is as filed and may therefore contain an incomplete specification.

METHOD TO PREPARE DYE-BASED REFERENCE MATERIAL Abstract

Provided is a method of preparing dye-based reference materials useful for calibrating or qualifying instrument systems that are diagnostic spectroscopically for hemoglobin and CO-ox fractions. The dye-based reference materials are non-proteinaceous and may be interpreted by CO-oximeter instrument systems as providing a spectrum substantially equivalent to the spectrum of blood,

- particularly with regard to hemoglobin and hemoglobin CO-ox 10 fractions. Dye-based reference materials may be formulated to provide CO-ox fractions approximating those seen in a clinical setting, including normal physiological CO-ox The dye-based reference materials may be fractions.
 - combined with other ingredients thus making the materials useful as references for pH and blood gas instruments and/or electrolyte instruments in addition to CO-oximeter instrument systems.

"EXPRESS MAIL" Mailing Label No. 18579534637

Date of Deprest:

i hereby certify that this poper or fee is being deposited with the United States Postal Service "Property Mail Post Office to Authorists" service under 37 CFR 1.10 on the date indicated show and Arthressed to the Consulections of Patrice and

Textionede: Westington/D.C.

20

15

That which is claimed is:

- 1. A method to prepare a reference material for CO-oximeter instrument systems is provided, said method comprising (a) selecting a red dye having a main absorption maxima at a wavelength region between about 560 to about 580 nm to provide a basic spectrum; (b) modifying said main absorption maxima of said red dye by using a dye having an absorption maxima at a wavelength region between about 350 to about 450 nm; and (c) establishing a new relative baseline of said spectrum at wavelengths greater than about 640 nm and providing a sufficient absorbance at about 630 nm to provide a positive methemoglobin CO-ox fraction value using one or more dyes.
- 2. A method according to claim 1 wherein said red dye has an absorption band width of less than about 50 nm and an absorption of near about zero at a wavelength of greater than about 610 nm.
- 3. A method according to claim 1 wherein said dye
 20 of step (b) is a yellow dye and said main absorption maxima

is shifted to a wavelength in the region of 570 to about 580 nm.

- 4. A method according to claim 1 wherein said step(c) is accomplished using two dyes, one of said two dyes is a blue dye having an absorption peak at about 625 to about 640 nm, and the second of said two dyes is a dye with an absorbance maxima positioned at a wavelength of from about 700 to about 780 nm and an absorption band width of from about 150 to about 200 nm.
- 5. A method according to claim 4 wherein said reference provides normal physiological CO-ox fractions of: oxyhemoglobin 94-98 %, carboxyhemoglobin 0 about 1 % methemoglobin 0 to about 1.5 %, and reduced hemoglobin from about 1 to about 5%.
- dye having a main absorption maxima at a wavelength between about 560 to about 580 nm; (b) a second dye having an absorption maxima at a wavelength between about 350 to about 450 nm; and (c) a third dye having an absorption maxima at a wavelength between about 640 nm; and (d) a fourth dye having an absorption maxima at a wavelength

between about 700 to about 780 nm and an absorption band width of from about 150 to about 200 nm.

- 7. A material according to claim 6 wherein each of said dyes present in a sufficient amount to provide said material with clinically meaningful CO-ox fractions.
- 8. A material according to claim 7 wherein said first dye is a red dye selected from the group consisting of sulforhodamine B, chlorophenol red, xylenol orange, and sulforhodamine 101; said second dye is a yellow dye selected from the group consisting of mordant yellow 7, tartrazine, orange G, hydroxypyrenetrisulfonic acid, mordant yellow 10, combinations thereof; said third dye is selected from the group consisting of brilliant blue FC (FD&C blue 1), erioglaucine, lissamine, alphazurine A, patent blue violet, patent blue VF, hydroxy naphthol blue, and combinations thereof; and said fourth dye is selected from the group consisting of IR 125, naphthol green B, and combinations thereof.
- 9. A material according to claim 8 wherein said
 20 first dye is sulforhodamine B; said second dye is mordant

15

yellow 7; said third dye is patent blue violet; and said fourth dye is naphthol green B.

- 10. A material according to claim 8 wherein said first dye is sulforhodamine B; said second dye is
- 5 tartrazine; said third dye is patent blue violet; and said fourth dye is naphthol green B.
 - 11. A material according to claim 8 wherein said first dye is sulforhodamine B; said second dye is mordant yellow 10; and said third dye is patent blue VF; and said fourth dye is naphthol green B.

10

- 12. A material according to claim 8 wherein said first dye is sulforhodamine B; said second dye is mordant yellow 7; and said third dye is FD&C blue 1; and said fourth dye is naphthol green B.
- 13. A material according to claim 8 wherein the first dye is used in an amount ranging from about 0.55 mM to about 4 mM; the second dye is used in an amount ranging from about 1 mM to about 10 mM; the third dye is used in an amount ranging from about 0.01 mM to 0.3 mM; and the fourth dye is used in an amount ranging from about 0.10 mM to 0.3 mM; and the fourth

14. A material according to claim 8 comprising the following:

COMPOUND	CONCENTRATIONS
NaCl	110 to 160 mM
KCl	2.5 to 7.5 mM
CaCl ₂	2 to 4 mM
HEPES	35 to 40 mM
TRIS.HCl	15 to 40 mM
NaOH	20 to 50 mM
EDTA	1 to 1.7 mM
NaHCO ₃	15 to 20 mM
LiCl	10 to 25 mM
Glucose	0.5 to 10 g/L
Li lactate	0.5 to 14 mM
Sulforhodamine B	0.55 to 1.3 g/L
Mordant Yellow 7	1.4 to 3 g/L
Patent Blue Violet	0.07 to 0.08 g/L
Naphthol Green B	0.4 to 1.1 g/L

15. A material according to claim 14 wherein said

5 reference liquid provides clinical CO-ox fractions of: about
60 % to about 98 % oxyhemoglobin; from 0 to about 20 %

carboxyhemoglobin; from 0 to about 20 % methemoglobin; and from 0 to about 20 % reduced hemoglobin.

- 16. A material according to claim 14 wherein said reference liquid provides physiological CO-ox fractions of: oxyhemoglobin 94-98 %, carboxyhemoglobin 0 about 1 %, methemoglobin 0 to about 1.5 %, and reduced hemoglobin from about 1 to about 5%.
- instrument systems comprising: (1) subjecting said
 instrument to a quality control standard material, said
 material prepared according to a method comprising (a)
 selecting a red dye having a main absorption maxima at a
 wavelength region between about 560 to about 580 nm to
 provide a basic spectrum; (b) modifying said main

 absorption maxima of said red dye by using a dye having an
 absorption maxima at a wavelength region between about 350
 to about 450 nm; and (c) establishing a new relative
 baseline of said spectrum at wavelengths greater than about
 640 nm and providing a sufficient absorbance at about 630 nm
 to provide a positive methemoglobin CO-ox fraction value
 using one or more dyes, wherein said dyes are used in a

sufficient amount to provide said control with a predetermined level of CO-ox fractions; (2) obtaining an instrument measurement of CO-ox fractions of said control; and (3) comparing said instrument measurement of said CO-ox fractions of said control with said predetermined level of CO-ox fractions to check the accuracy of said instrument.

A method for quality control of CO-oximeter instrument systems comprising: (1) subjecting said instrument to a quality control standard material comprising (a) a first dye having a main absorption maxima at a 10 wavelength between about 560 to about 580 nm; (b) a second dye having an absorption maxima at a wavelength between about 350 to about 450 nm; and (c) a third dye having an absorption maxima at a wavelength between about 625 and about 640 nm; and (d) a fourth dye having an absorption 15 maxima at a wavelength between about 700 to about 780 nm and an absorption band width of from about 150 to about 200 nm; wherein said dyes are used in a sufficient amount to provide said control with a pre-determined level of CO-ox fractions; (2) obtaining an instrument measurement of CO-ox fractions of said control; and (3) comparing said instrument

measurement of said CO-ox fractions of said control with said predetermined level of CO-ox fractions to check the accuracy of said instrument.

- 19. A method according to claim 18 wherein three levels of quality control standards are prepared.
 - 20. A method according to claim 19 wherein said three levels of standards comprising the following:

Dye	Level a	Level b	T 1
Sulforhodamine B			Level c
	0.98 mM	1.95 mM	2.27 mM
Mordant Yellow 7	6.65	7.63	3.89
Patent Blue Violet	0.14	0.11	0.14
Naphthol Green B	0	1.10	0.48

21. A method according to claim 19 wherein said three levels of standards comprising the following:

Dye	Level a	Level b	Level c
Sulforhodamine B	1.27 mM	2.53 mM	3.22 mM
Tartrazine	1.34	1.87	1.60
Patent Blue Violet	0.11	0.17	0.25
Naphthol Green B	0.98	1.96	1.64

22. A method according to claim 19 wherein said three levels of standards comprising the following:

Dye ·	Level a	Level b	Level c
Sulforhodamine B	0.93 mM	2.14 mM	2.37 mM
Mordant Yellow 10	8.95	9.24	4.91
patent blue VF	0.12	0.13	0.08
Naphthol Green B	0	1.33	0.81

- 23. A method according to claim 19 wherein said
- 5 three levels of standards comprising the following:

Dye	Level a	Level b	Level c
Sulforhodamine B	1.17 mM	1.97 mM	2.19 mM
Mordant Yellow 7	5.33	6.14	2.62
FD&C Blue 1	0.03	0.04	0.03
Naphthol Green B	0.97	2.30	0.14

METHOD TO PREPARE DYE-BASED

REFERENCE MATERIAL

Field of Invention

This invention relates to a method of preparing dye-based reference materials useful for calibrating or qualifying instrument systems that are diagnostic spectroscopically for hemoglobin and CO-ox fractions. The reference materials may also be used in instrument systems that have, in addition to capability to measure CO-ox fractions, sensors for the measurement of blood pH, gas and other blood analytes including electrolyte concentrations and metabolite concentrations.

Background of the Invention

Improvements in instrumentation have made the

determination of blood pH, gas, electrolytes and CO-ox

fractions relatively routine in clinical laboratories.

Typically, pH and blood gas instruments measure blood pH,

pCO₂, and pO₂. CO-oximeter instruments typically measure

the total hemoglobin concentration (THb); the hemoglobin

fractions such as oxyhemoglobin (O₂Hb); methemoglobin

(MetHb); carboxyhemoglobin (COHb); and reduced hemoglobin

(HHb) (collectively referred to herein as "CO-ox fractions"). Electrolyte instruments measure any number of blood electrolytes, including sodium, potassium, lithium, calcium, and so on. Instrument systems currently available may combine the measurement of blood pH, gases, electrolytes, various metabolites, and CO-ox fractions in one instrument for a comprehensive testing of the properties of blood, as is particularly useful in respiratory and pulmonary ailments. Vigorous therapeutic treatment is often dictated by such test results.

Quantitative determination of the various CO-ox fractions in clinical settings is desirable as CO-ox fractions relate to the loading of oxygen onto the hemoglobin of red blood cells circulating through the pulmonary capillaries. The actual amount of oxygen loaded onto the hemoglobin is determined not only by the concentration of total hemoglobin (THb), but also by the amount of non-oxygen-binding derivatives of hemoglobin such as carboxyhemoglobin (COHb) and methemoglobin (MetHb).

20 Reduced hemoglobin (HHb) is an unoxygenated form of normal hemoglobin and elevations in the arterial fractional HHb

indicate that lesser amounts of oxygen have been bound as a result of ventilation and/or perfusion defects.

Blood, hemoglobin and hemoglobin CO-ox fractions absorb visible light. A normal blood spectrum has a main absorption peak at 578 nm and decreases rapidly close to zero at wavelengths greater than about 610 nm as shown in Figure 1. The second absorption peak of blood is at 542 nm. Absorbance maxima of hemoglobin derivatives are oxyhemoglobin, 541, 568 - 572 nm; reduced hemoglobin, 555 nm; carboxyhemoglobin, 537, 568 - 572 nm; and methemoglobin 540, 578, 630 nm.

10

Generally, optical type of CO-oximeters measure
the absorbance of the blood sample at multiple wavelengths
on the spectrum. Ultimately, based on the known CO-ox
fraction absorption wavelengths regions, CO-oximeters
analyze blood samples by the collection of absorption data
at specific wavelengths. The data are then typically
recorded and a process called multicomponent analysis is
used to simultaneously calculate the concentrations of the
each of hemoglobin CO-ox fraction present in the blood
sample.

CO-oximeter instruments are typically designed to measure CO-ox fractions with values broader than the ranges observed in patient samples. For example, instruments may have the capability to measure CO-ox fractions in the defined instrument ranges of: about 4 to about 25 g/dL of THb, about 30 % to about 98 % oxyhemoglobin (O2Hb); 0 to about 50 % carboxyhemoglobin (COHb); 0 to about 40 % methemoglobin (MetHb); and 0 to about 50 % reduced hemoglobin (HHb), with all CO-ox fraction percentages herein 10 based on the total amount of hemoglobin. Within the range of instrument capability is a clinically meaningful CO-ox fraction range and a normal physiological CO-ox fraction The clinically meaningful range is defined herein as: total hemoglobin 8 to 20 g/dL, 60 to 98 % oxyhemoglobin 15 (O2Hb); 0 to 20 % carboxyhemoglobin (COHb); 0 to about 20 % methemoglobin (MetHb); and 0 to about 20 % reduced hemoglobin (HHb). The normal physiological range is defined herein as: total hemoglobin 14 to 17 g/dL for men and 12 to 15 g/dL for women, oxyhemoglobin 94 to 98 %,

20 carboxyhemoglobin 0 to about 1 % for nonsmoker,

methemoglobin 0 to about 1.5 %, and reduced hemoglobin from about 1 to about 5%.

Reference materials generally function to validate the performance of a diagnostic instrument. For CO-oximeter instrument systems, ideal quality control standard materials are formulated to provide pre-determined CO-ox fraction values not only within the broad instrument capability range, but also within the normal physiological range.

oximeter quality control standard materials. The first type are aqueous dye-based materials, where dyes are used in an attempt to match the spectrum of blood. The second type are blood-based materials, where the presence blood allows for the direct match of the spectrum of blood. Both types of materials have been associated with a variety of problems as discussed below.

In developing dye-based quality control standard materials for CO-ox instrument systems, to more closely approximate the spectrum of blood, combinations of dyes have been used. The combinations of dyes have been used because no single synthetic dye has a spectrum sufficiently similar

20

to the absorption bands of a blood spectrum. Because the spectrum of blood has multiple distinctive bands of absorption, it is challenging to prepare a quality control standard material to mimic these characteristics. Although one dye might contribute an absorption characteristic that is present in the normal blood spectrum, it may also present other absorption characteristics that are dissimilar to blood at other portions of the spectrum.

Prior art methods teach dye combinations that

only partially simulate blood's visible spectrum.

Consequently, the CO-ox fraction values of the prior art
quality control standard materials are often not clinically
meaningful. For example, U.S. Patent 4,843,013 teaches a
combination of dyes for a CO-ox quality control standard,

however, as shown in Cols. 6 and 7 of U.S. Patent 4,843,013,
the described quality control standard provides negative
values for some of the CO-ox fractions. A negative value
for a CO-ox fraction would never appear in a blood sample
and thus these control standards have limited usefulness in
qualifying CO-oximeters in a clinical setting. There has
been a long-felt commercial need for dye-based quality

control materials that may be formulated in a predictable manner to provide pre-determined clinically meaningful CO-ox fractions, particularly in the normal physiological range.

Blood-based reference materials provide

clinically meaningful CO-ox fraction values, however,

numerous limitations have been associated with using blood

in the quality control standard materials. For example, a

significant problem encountered by users of the

proteinaceous blood-based materials is that blood-based

materials are very susceptible to bacterial contamination

and have a limited shelf life. Consequently, in most cases,

users of blood-based materials must refrigerate the products

during storage. Additionally, blood-based reference

materials are generally classified as biohazardous

materials, thus requiring the user to take additional safety

precautions.

There is a need to provide non-proteinaceous dyebased reference materials that provide clinically and physiologically meaningful CO-ox fraction ranges.

SUMMARY OF THE INVENTION

10

15

According to the invention, a method to prepare a reference material for CO-oximeter instrument systems is provided, said method comprising (a) selecting a red dye having a main absorption maxima at a wavelength region between about 560 to about 580 nm to provide a basic spectrum; (b) modifying said main absorption maxima of said red dye by using a dye having an absorption maxima at a wavelength region between about 350 to about 450 nm; and (c) establishing a new relative baseline of said spectrum at wavelengths greater than about 640 nm and providing a sufficient absorbance at about 630 nm to provide a positive methemoglobin CO-ox fraction using one or more dyes.

Further provided is a reference material prepared according to said method, wherein said dyes are used in sufficient amounts to provide a reference material with CO-ox fractions falling within a range covering instrument capability. The method may also be used to prepare reference materials having CO-ox fractions with values falling within the defined clinical range or within the defined normal physiological range.

Also provided in one embodiment is a reference material comprising (a) a first dye having a main absorption maxima at a wavelength between about 560 to about 580 nm;

(b) a second dye having an absorption maxima at a wavelength between about 350 to about 450 nm; and (c) a third dye having an absorption maxima at a wavelength between about 625 and about 640 nm; and (d) a fourth dye having an absorption maxima at a wavelength between about 780 nm and an absorption band width of from about 150 to about 200 nm. The amounts of the dyes used in the materials may be adjusted to provide quality control materials with desired CO-ox fraction values.

Further provided is a method for quality control
of CO-oximeter instrument systems comprising: (1)

15 subjecting said instrument to a quality control standard
material comprising (a) a first dye having a main absorption
maxima at a wavelength between about 560 to about 580 nm;

(b) a second dye having an absorption maxima at a wavelength
between about 350 to about 450 nm; and (c) a third dye

20 having an absorption maxima at a wavelength between about
625 and about 640 nm; and (d) a fourth dye having an

absorption maxima at a wavelength between about 700 to about 780 nm and an absorption band width of from about 150 to about 200 nm; wherein said dyes are used in a sufficient amount to provide said control with a pre-determined level of CO-ox fractions; (2) obtaining an instrument measurement of CO-ox fractions of said control; and (3) comparing said instrument measurement of said CO-ox fractions of said control with said predetermined level of CO-ox fractions to check the accuracy of said instrument.

The dye-based reference materials are nonproteinaceous and may be interpreted by CO-oximeter
instrument systems as providing a spectrum substantially
equivalent to the spectrum of blood, particularly with
regard to total hemoglobin and hemoglobin CO-ox fractions.

10

Using the inventive method of preparing the reference materials, dye-based reference materials may be formulated to provide CO-ox fractions approximating those seen in a clinical setting, including normal physiological values. These dye-based reference materials may also be combined with other ingredients which make the materials useful as reference materials for pH and blood gas instruments and/or

electrolyte instruments in addition to CO-oximeter instruments.

BRIEF DESCRIPTION OF DRAWINGS

FIG. 1 shows the spectrum of blood.

FIG. 2 shows the spectrums of single dyes and a dye combination that does not provide clinically meaningful CO-ox fractions. Spectrum 1 is taken from a sulforhodamine B solution. Spectrum 2 is taken from a sulforhodamine 101 solution. Spectrum 3 is taken from a solution containing 1 g/L sulforhodamine B and 0.1 g/L sulforhodamine 101.

10

FIG. 3 illustrates spectroscopically how the spectrum of sulforhodamine B is modified using the inventive method of formulation. "SRB" is an abbreviation for sulforhodamine B; "MY7" is an abbreviation for mordant yellow 7; "PBV" is an abbreviation for patent blue violet; and "NGB" is an abbreviation for naphthol green B.

FIGs. 4-7 show the spectra of the three levels of formulations shown in Examples 1-4, respectively.

DETAILED DESCRIPTION OF INVENTION

The present invention provides a method to formulate dye-based reference materials which mimic the

spectrum of human blood when the reference materials are utilized in optical types of CO-oximeter instrument systems. The formulating method is described herein in distinct steps for purposes of clarity only. It should be understood that the steps set forth may be carried out in any order and may also be combined and carried out simultaneously. Additionally, the designation of first dye, second dye, and so on, is not intended to imply a specific order of dyes. The types of dyes selected, the amount of dyes utilized and the concentration of dyes described in these steps may be 10 adjusted until the desired level of hemoglobin CO-ox fractions are obtained, as preferably determined by taking a reading from the specific CO-oximeter instrument that will be used with the reference material. All measurements of CO-ox fractions set forth herein were taken on Ciba Corning M270 CO-oximeters (as available from Ciba Corning Diagnostics Corp., Medfield, Mass. USA). All measurements of blood gas and electrolytes set forth herein were taken on Ciba Corning M288 Blood Gas Analyzers.

20 According to the invention, in step (a) a red dye is selected to provide a basic spectrum (also referred to

herein as the first dye). In selecting the red dye, the general characteristics of the spectrum of normal blood should be considered. Normal blood has a spectrum with a main absorption peak at 578 nm which decreases rapidly close to zero at a wavelength greater than about 610 nm as shown in Figure 1. The second absorption peak of blood is at approximately 542 nm. The absorbance maxima of the CO-ox fraction oxyhemoglobin is at the approximate wavelengths of 541 nm and 568 - 572 nm. As normal arterial blood has an oxyhemoglobin fraction greater than about 95%, the dye-based formulation should have a red dye with a spectrum similar to at least the general absorption features of oxyhemoglobin. The preferred red dyes in the invention have an absorption peak in the wavelength range from about 560 nm to about 580 nm. Further, particularly preferred red dyes exhibit minimal absorbance at a wavelength of greater than about 610 nm. Red dyes, such as, for example, sulforhodamine B, chlorophenol red, xylenol orange, sulforhodamine 101, etc., are currently preferred. preferably employed as the red dye is sulforhodamine B, which has an absorption maximum at approximately 566 nm in

aqueous solution, a fairly narrow absorption bandwidth, and an absorbance which decreases to zero at about 620 nm, as shown as spectrum 1 in Figure 2.

The single red dye does not exhibit a spectrum sufficiently similar to that of blood for purposes of obtaining clinically meaningful CO-ox fractions, as exemplified by the CO-ox fractions measured from 1.2 g/L of sulforhodamine B: THb 14 g/dL, O₂Hb -16%, COHb 105%, MetHb -2.5% and HHb 13.5%. As these values indicate, negative CO-ox fractions are obtained for the oxyhemoglobin and the methemoglobin.

10

One approach to alter the absorption peak of the red dye might be to introduce another red dye. In this approach, to shift the absorption peak to a longer wavelength, another red dye with absorption maximum wavelength greater than that of the first red dye could be utilized. Alternatively, to shift absorption peak to a shorter wavelength, another red dye with absorption maximum wavelength less than that of the first red dye could be utilized. A combination of the two red dyes would be predicted to result in a new overlapped maximum that would

complement the required absorption features. However, as shown in Figure 2, sulforhodamine B in aqueous solution provides a spectrum that does not have enough strength of absorbance at 570 to 585 nm because the absorption peak of the sulforhodamine B is located at a shorter wavelength than that of blood. Sulforhodamine 101 has an absorption peak at 587 nm. A combination of sulforhodamine B and sulforhodamine 101 shifts the absorption peak of sulforhodamine B and enhances the absorbance at 570 $nm\ to$ 10 In Figure 2, spectrum 1 is sulforhodamine B, spectrum 2 is sulforhodamine 101, and spectrum 3 is the combination of 1 g/L sulforhodamine B and 0.1 g/L sulforhodamine 101. As shown in Figure 2, absorbance of the two red dye combination at 570 to 585 nm is enhanced and results in the change of CO-ox fractions from negative values to positive values, e.g., 14 g/dL THb, 16% O_2 Hb, 1% MetHb, 14% COHb and 69% HHb. The addition of the second red dye (sulforhodamine 101) introduces an undesired high absorbance near wavelength range of 590 nm to 605 nm, thus not providing an acceptable method of formulating references

having clinically meaningful and/or normal physiological CO-ox fraction values.

In step (b), modification and manipulation of the absorption band of the red dye is conducted where the absorption peak of the red dye is shifted to a longer wavelength to more closely approximate that of blood. Upon step (b), the absorption maxima of the red dye is shifted to fall between the wavelengths of approximately 570 nm to about 585 nm. It is preferred that the shift of the absorption maxima does not inadvertently increase the absorption in the wavelength region of 590 to 605 nm. It has been found that dye(s) having an absorption maxima at wavelengths of about 350 to 450 nm are useful in step (b) because the dye(s) increase the absorbance of the spectrum at the ultraviolet region and force the absorption band of the red dye to shift toward the longer wavelength.

10

Preferably in step (b) is accomplished using a yellow dye (or a combination thereof) including, for example, mordant yellow 7, tartrazine, orange G, hydroxypyrenetrisulfonic acid, mordant yellow 10, combinations thereof, and so on. Particularly preferred is

mordant yellow 7 dye (especially when sulforhodamine B is selected as the red dye).

Figure 3 illustrates the changes of the spectrum

of sulforhodamine B after addition of mordant yellow 7.

5 Spectrum 1 in Figure 3 is the sulforhodamine B aqueous solution, and spectrum 2 is the sulforhodamine B aqueous solution containing mordant yellow 7. As shown, the original absorption peak of sulforhodamine B is shifted from 566 nm to 572 nm, with the edge of the absorption band at approximately 610 nm, and with little increase of absorbance

in the wavelength range of 590 to about 605 nm.

According to the present invention, in step (c) a new relative baseline of the spectrum is established at wavelengths greater than about 640 nm and sufficient

15 absorbance at about 630 nm is provided to yield a positive methemoglobin CO-ox fraction. Step (c) addresses that the absorbance at about 590 to about 605 nm of the dye or dye combination used in steps (a) and (b) is higher than that of normal blood. This step may be accomplished using one or more dyes. With respect to the new lifted baseline, the original absorbance can be normalized, with the normalized

absorbance reduced and smaller than the original absorbance. Further, the original absorbance at about 590 to about 605 nm, although having a similar absolute absorbance value, is relatively reduced upon the establishment of the new lifted baseline. CO-oximeters take the lifted baseline from the CO-oximeters interpret the dye spectrum, and calculate the CO-ox fractions therefrom. Upon step (c), the inventive reference material will read as providing CO-ox fraction values within the instrument capability and clinically meaningful ranges, and optionally, within the normal physiological range, as described in more detail hereinafter.

10

One group of dyes particularly useful in step (c) are designated herein as the third dye(s). These dyes have an absorption maxima at about 625 to about 640 nm (more preferably about 630 nm) and have only minimum impact on the absorbance at wavelengths shorter than about 600 nm.

Preferably, the third dye is a blue dye selected from brilliant blue FC (FD&C blue 1), erioglaucine, lissamine, alphazurine A, patent blue violet, hydroxy naphthol blue, patent blue VF, combinations thereof, and equivalents

thereof (with patent blue violet, patent blue VF, and alphazurine more preferred). Patent blue violet has an absorption maxima at about 639 nm and is most preferred because it provides the absorbance not only at 630 nm, but also absorbance to some extent at a wavelength greater than 640 nm, thus generally accomplishing both functions of step (c). Adding the patent blue violet into the sulforhodamine B and yellow mordant 7 combination pushes the CO-ox fractions to the following: O₂Hb 80%, MetHb 15%, COHb 2% and HHb 3 %, thus placing the reference material within the values set forth for instrument capability and clinically meaningful CO-ox fractions.

A second group of dyes that may be used in step

(c) are designated as the fourth dye(s). These dyes have an
main absorbance maxima positioned at a wavelength of from
about 700 to about 780 nm and an absorption band width of
from about 150 to about 200 nm. The tail of the absorption
band of the dye (or dyes) of step (c) preferably lifts the
baseline of the absorbance spectrum with minimal effect on
absorbance at 500 to 605 nm. Preferred dyes falling within
this description may be selected from any number of dyes

including IR 125, naphthol green B, and combinations
thereof, and so on. Most preferred is naphthol green B,
having an absorption maxima at approximately 716 nm in an
aqueous solution, and considered particularly effective when
combined with mordant yellow 7, patent violet blue and
sulforhodamine B.

For step (c) the combination of the third and fourth dyes allow the reference material containing the first and second dyes to yield CO-ox fraction values falling with the normal, physiological range. Figure 3 exemplifies the principles of the present method of formulation. The sulforhodamine B (SRB) is the red dye that provides a basic The addition of mordant yellow 7 (MY7) to the sulforhodamine B (SRB) modifies the spectrum by tilting the absorption band of the sulforhodamine B towards a longer wavelength and shifts the absorption maxima from about 566 nm to about 572 nm (SRB + MY7), to more closely approximate the main absorption spectrum of blood. Patent blue violet (PBV) is added for purposes of providing a spectrum absorption for methemoglobin at approximately 630 nm (SRB + MY7 + PBV) and also slightly lifts the baseline spectrum.

10

15

20

The addition of the naphthol green B (NGB) further lifts the spectrum baseline to the right of the main absorption The ultimate result is shown by Spectrum "SRB + MY7 + PBV + NGB". In this formulation the following amounts of dyes were used: about 1.15 g/L sulforhodamine B; about 2.4 g/L mordant yellow 7; about 0.08 g/L patent blue violet; and about 1 g/L naphthol green B. When this combination was measured on a Ciba Corning M270 CO-oximeter the following CO-ox fraction ranges (which are within the normal physiological range) were provided: THb 14 g/dL, O2Hb 96 %, COHb 1%, MetHb 0.8% and HHb 2.2 %. Providing a reference material having an oxyhemoglobin value greater than about 60 % has been met with limited success in the past, however, the present invention provides a method to easily yield this 15 objective.

Any suitable dyes may be used in the formulation of the reference materials with either aqueous or nonaqueous solvents. Water soluble dyes are particularly well-suited, including acid dyes, basic dyes, direct dyes, and so on, and equivalents thereof. The dye composition may be prepared as a dry material for ease of storage and

If prepared as a dry composition, prior to usage packaging. the composition may be prepared as a solution using a suitable liquid, including water and various organic solvents, or mixtures thereof and so on, by techniques well known to those skilled in the art. It is particularly preferred that compatible dyes are used, with a particularly preferred embodiment utilizing anionic dyes. Although the method of formulating may be accomplished using various amounts of dyes, a particularly preferred composition employs a total dye concentration of from about 10 0.1 to about 10 mM. Further, the stability of the reference materials will be increased when high purity dyes, which are either commercially available or purified, using conventional methods known to those skilled in the art, are 15 used in formulation.

The reference materials prepared according to the invention may be utilized on various types of optical CO-oximeters (or instruments incorporating a CO-oximeter function). Because various types of CO-oximeters may be designed and constructed to measure the absorbance of blood at different set of wavelengths, the reference materials

used for these CO-oximeters will typically need dye concentration adjustment for this variation, where the adjustment may be accomplished by those skilled in the art employing the principles of this invention.

The reference materials may be prepared as calibration materials to simultaneously calibrate instrument systems that include sensors for blood gas, electrolyte, and various metabolites in addition to the CO-oximeter capability. Further, the reference materials are particularly well-suited for preparation as quality control 10 materials for instrument systems that include a CO-oximeter capability, in addition to various other sensors, as described herein. For CO-oximeter instrument systems, ideally, the CO-ox fraction accuracy is generally checked both in the broader instrument capability CO-ox fraction 15 range as well as in the more narrow normal physiological range (as defined herein). More preferably, the CO-oximeter instrument system is checked for performance accuracy on at least three levels. These three reference materials are preferably formulated to provide pre-determined CO-ox fraction values for the total hemoglobin as well as various

hemoglobin CO-ox fractions outside of and within normal physiological ranges.

Particularly preferred compositions provide a commercially acceptable stability. The stability may also extend to high temperature stability such that the composition may be autoclaved for purposes of facilitating sterilization, if desired. Additionally, the dye formulations may be prepared such that they are stable to exposure to light.

Appropriate ingredients that be included in the dye formulation preferably are those ingredients that do not interfere with the sensors of the instruments such that the reference materials may be used to provide calibration or quality control of different types of sensors

simultaneously. A preferred reference prepared using the dye formulation of this invention may be prepared by combining in the appropriate relative proportions, buffer(s), dyes, sources of analytes (typically salts), and other standard chemicals commonly used in reference

materials including preservatives and surfactants, and so on, as well known in the art. Buffer materials that may be

used, particularly when preparing a reference standard for simultaneous usage on pH blood gas and electrolyte sensor system, are preferably selected for having a pKa close to the desired working pH. Particularly useful buffer materials are zwitterion buffers that are compatible with the selected dyes, including, for example, N-2hyroxyethylpiperazine-N'-2-ethanesulfonic acid (HEPES), 3-(N-morpholinio) propanesulfonic acid (MOPS), Tris-(Hydroxymethyl) aminomethane (TRIS), and so on, as are well known to those skilled in the art. A combination of the salts such as, for example, NaCl, NaOH, KCl, LiCl, NaHCO3, Na₂SO₄, CaCl₂, glucose, lithium lactate, and equivalents thereof, can be used to provide the desired concentration of electrolytes, as is well known to those skilled in the art. A stable ionized calcium concentration may be provided by using calcium chelating agents, such as, ethylenediaminetetraacetic acid (EDTA), Nhydroxyethyliminodiacetic acid (HIDA), nitrilotriacetic acid (NTA), and citric acid. Further, when providing a reference for a blood gas measurement, the desired pCO2 and pO2 ranges 20

may be reached by tonometering to an equilibrium at

approximately 25° C using gas combination with proper oxygen, carbon dioxide, and nitrogen contents, as well within the skill of those trained in the art.

A particularly useful reference material includes the above-described dyes providing normal physiological Co-ox fractions while in combination with an aqueous solution of known quantities of pH, pO₂, pCO₂, Na, K, Cl, and ionized Ca (iCa), glucose, and lactate (preferably in amounts that typically fall within the range of a blood sample). After said combination, the reference material may be divided into aliquots which are placed in sealed receptacles. Thereafter the receptacles may be flushed with tonometry gas, and the receptacle sealed. These references may be used for verifying the accuracy and reliability of instrument systems having multiple sensors for measuring blood pH, gas, CO-ox fractions and various electrolyte and metabolite levels. A particularly preferred embodiment that is described in TABLE A below.

10

TABLE A			
COMPOUND	APPROXIMATE CONCENTRATIONS		
NaCl	110 to 160 mM		
KCl	2.5 to 7.5 mM		
CaCl ₂	2 to 4 mM		
HEPES	35 to 40 mM		
TRIS.HCl	15 to 40 mM		
NaOH	20 to 50 mM		
EDTA	1 to 1.7 mM		
NaHCO ₃	15 to 20 mM		
LiCl	10 to 25 mM		
Glucose	0.5 to 10 g/L		
Li lactate	0.5 to 14 mM		
Sulforhodamine B	0.55 to 1.3 g/L		
Mordant Yellow 7	1.4 to 3 g/L		
Patent Blue Violet	0.07 to 0.08 g/L		
Naphthol Green B	0.4 to 1.1 g/L		

To prepare clinically meaningful reference materials, preferably the first dye is used in an amount ranging from about 0.55 mM to about 4 mM; the second dye is used in an amount ranging from about 1 mM to about 10 mM; the third dye is used in an amount ranging from about 0.01 mM to 0.3 mM; and the fourth dye is used in an amount ranging from 0 to about 2.5 mM.

It is to be understood that various modifications to the invention will be apparent to and can readily be made by those skilled in the art, given the disclosure herein, without departing from the scope and materials of this invention. It is noted that the following examples given herein are intended to illustrate and not to limit the invention thereto.

EXAMPLES

Example 1

10 A three-level quality control standard material was prepared with varying levels of pH, pCO₂, pO₂, sodium (Na), potassium (K), chloride (Cl), ionized calcium (iCa), and three levels of total hemoglobin (THb), oxyhemoglobin (O₂Hb), carboxyhemoglobin (COHb), methemoglobin (MetHb), and 15 reduced hemoglobin (HHb), as shown in TABLES I and II. The three levels prepared represent clinically significant ranges of acidosis (level a), normal or physiological (level b), and alkalosis (level c).

More particularly, three bulk solutions containing
the required salts were prepared in an aqueous solution, as
shown below in TABLE I.

TABLE I				
BC	BULK SOLUTIONS			
COMPOUND	Level a	Level b	Level c	
NaCl	55.5 mM	60 mM	65 mM	
KCl	3.20	4.85	710	
CaCl ₂	3.81	2.80	2.13	
NaOH	29.5	38	34.3	
LiCl	21.5	12.9		
EDTA	1.63	1.20	1.14	
HEPES	40	40	40	
TRIS.HCl	35.5	15		
TRIS base	10			
NaHCO ₃	19.5	23.4	19.5	

To the bulk solutions described in TABLE I, three levels of dye formulations were added, as shown in TABLE II.

TABLE II DYE FORMULATIONS				
Dye Level a Level b Level				
Sulforhodamine B	0.98 mM	1.95 mM	2.27 mM	
Mordant Yellow 7	6.65	7.63	3.89	
Patent Blue Violet	0.14	0.11	0.14	
Naphthol Green B		1.10	0.48	

After dyes dissolved in the bulk solution, the solutions were then tonometered at 25° C to an equilibrium using the following gas mixtures.

5

TABLE III					
	GAS CONCENTRATIONS (%)				
Gas	s Level a Level b Level c				
CO ₂	6.10	3.15	1.77		
02	5.42	11.60	17.45		
N ₂	88.48	85.25	80.78		

After tonometry equilibrium was established, the pH, pO₂, and pCO₂ of the solutions remained constant even after a prolonged period of tonometry. Aliquots of the tonometered solutions were then transferred into glass ampoules while purging with the same gas mixture used in the solution tonometry. The filled ampoules were then flamesealed.

The three levels of quality control standard having the combinations listed in TABLES I, II, and III were then evaluated by two types of instruments: ANALYZER 1 = Ciba Corning M288 blood gas analyzer and ANALYZER 2 = Ciba Corning M270 CO-oximeter. Results are summarized in TABLE IV.

TABLE IV				
Evaluations of Reference Standards				
ANALYTE	LEVEL a	LEVEL b	LEVEL C	
ANALYZER 1				
(Blood Gas)				
р́Н	7.15 ± 0.02	7.4 ± 0.02	7.6 ± 0.02	
pCO₂ mm Hg	70 ± 2	40 ± 2	21 ± 2	
pO ₂ mm Hg	63 ± 2	100 ± 3	150 ± 5	
Na mM	115 ± 2	135 ± 2	155 ± 3	
K mM	2.9 ± 0.3	4.8 ± 0.3	7.1 ± 0.3	
Cl mM	120 ± 2	100 ± 2	80 ± 2	
iCa mM	1.6 ± 0.05	1.15 ± 0.05	0.65 ± 0.05	
ANALYZER 2				
(CO-oximeter)				
THb g/dL	8.0 ± 0.2	14 ± 0.2	18 ± 0.2	
O₂Hb %	78 ± 2	95 ± 3	60 ± 2	
COHb %	3.5 ± 1	2.5 ± 1	6 ± 1	
MetHb %	16 ± 2	1 ± 0.3	18 ± 2	
HHb %	2.5 ± 1	1.5 ± 0.7	16 ± 2	

FIG. 4 provides the absorbance spectra of the three dye-based quality control standard described above, where spectrum a represents level a; spectrum b represents a

level b; and $spectrum\ c$ represents level c, each of which were measured using a 0.1 mm cuvette.

EXAMPLE 2

A three level quality control standard was prepared by

5 the procedure of Example 1. To the bulk solutions described in TABLE I of Example 1 the dye formulations shown in TABLE V below were added.

TABLE V				
DYE FORMULATIONS				
Dye Level a Level b Level c				
Sulforhodamine B	1.27 mM	2.53 mM	3.22 mM	
Tartrazine	1.34	1.87	1.60	
Patent Blue Violet	0.11	0.17	0.25	
Naphthol Green B	0.98	1.96	1.64	

The formulated solutions were then tonometered at 25°

C to equilibrium using the same gas mixtures described in TABLE III of Example 1. The tonometered solutions were then transferred into glass ampoules while purging the same gas mixture used in the solution tonometry. The filled ampoules were flame-sealed. Those three levels of reference were measured in the same manner as described in Example 1. The measured ranges were similar to the data shown in Example 1.

FIG. 5 provides the absorbance spectra of the three level dye-based quality control standard described above, where spectrum a represents level a; spectrum b represents a level b; and spectrum c represents level c, each of which were measured using a 0.1 mm cuvette.

Example 3

A three level quality control standard was prepared by the same procedure of Example 1. The bulk solutions of TABLE I were combined with the dye formulations of TABLE VII below.

10

15

TABLE VII				
Dye Formulations Dye Level a Level b Level c				
Sulforhodamine B	0.93 mM	2.14 mM	2.37 mM	
Mordant Yellow 10	8.95	9.24	4.91	
patent blue VF	0.12	0.13	0.08	
Naphthol Green B		1.33	0.81	

The formulated solutions were tonometered and transferred into ampoules, sealed, and thereafter measured, as described in Example 1. The data obtained using this formulation were similar to those measured in Example 1.

FIG. 6 provides the absorbance spectra of the three level dye-based quality control standard described above,

where spectrum a represents level a; spectrum b represents a level b; and spectrum c represents level c, each of which were measured using a 0.1 mm cuvette.

Example 4

A three level quality control standard was prepared by the same procedure of Example 1. The bulk solutions of TABLE I were combined with the dye formulations of TABLE VIII below.

TABLE VIII Dye Formulations				
Dye Level a Level b Level c				
Sulforhodamine B	1.17 mM	1.97 mM	2.19 mM	
Mordant Yellow 7	5.33	6.14	2.62	
FD&C Blue 1	0.03	0.04	0.03	
Naphthol Green B	0.97	2.30	0.14	

10

15


5


The formulated solutions were tonometered and transferred into ampoules, sealed, and thereafter measured, as described in Example 1. The data obtained using this formulation were similar to those measured in Example 1.

FIG. 7 provides the absorbance spectra of the three level dye-based quality control standard described above, where spectrum a represents level a; spectrum b represents

a level b; and spectrum c represents level c, each of which were measured using a 0.1 mm cuvette.


FIG. 1






Gowling, Strathy & Henderson


FIG. 3




Gowling, Strathy & Henderson



Powling Strains & Renderson




Gowling, Strathy & Henderson



Gowling, Strathy & Henderson

FIG. 7



Frankling Sought to Story Troops